



# Product Safety Recall



## Do you own this product?

### 500ml Essential Oil Distillation Kit

Lab distillation equipment kit for distilling essential oil or water.

Kit includes 500ml distillation flask, pear-shaped separating funnel, alcohol lamp, serpentine condenser, metal stand, accessories kit, and clamp kit.

Dates sold: 27/03/2019 – 02/03/23



**Why the product is recalled:** The mesh plate the flask sits on may contain asbestos. There is a risk of exposure to asbestos fibres when using this product.

**Hazard:** Exposure to airborne asbestos fibres can cause asbestosis, lung cancer and mesothelioma.



1. Immediately stop using this product and wrap it in thick sturdy plastic or heavy-duty sturdy plastic bag where the seal cannot be broken. Vevor will contact consumers to arrange return and replacement of the mesh plate.
2. Send us an email or phone us if you are unable to return the product or have sold or gifted it to another person



For more information contact Vevor on [poingsutia@hotmail.com](mailto:poingsutia@hotmail.com) or by phone on +86 16601839140.



Purchased from: online trader [www.vevor.com](http://www.vevor.com)

PRA2023/19753

See [productsafety.gov.au](http://productsafety.gov.au) for  
Australian product recall information