



Product Safety Recall



Do you own this product?

**UEBE Domotherm Rapid
REF0830**

White-and-coloured thermometer with digital display screen and metal tip

Dates sold: 22 June 2022 – 2 February 2023

Why the product is recalled: The product does not comply with the [mandatory safety standard for products containing button/coin batteries](#).



The packaging does not display the required warning label and does not display treatment advice in case of ingestion.

Hazard: Risk of severe internal burn injuries, which can result in serious injury or death if young children gain access to button batteries and ingest them.

Button batteries also pose a choking hazard to young children.



1. Consumers will be contacted by NaturalCycles by email to arrange a replacement thermometer.
2. Send us an email if you are unable to return the product or have sold or gifted it to another person



Ask our staff to find out more about the recall.

support@naturalcycles.com



Purchased from: [NaturalCycles](#)

PRA 2023/19786

See [productsafety.gov.au](https://www.productsafety.gov.au) for
Australian product recall information